

ACIS Verification Summary Report

| | |
|-----------------------|---|
| Specification: | AXAF Observatory to Science Instrument ICD (IF1-20) |
|-----------------------|---|

| | |
|----------------------------------|--|
| Requirement Number/Title: | 3.9.4.3 Nuclear Sources (VRSD 3.9.4.3-2) |
|----------------------------------|--|

Requirement Statement: Handling and control of flight equipment containing radioactive sources shall be in accordance with the AXAF safety plan, SA03.

| | |
|-----------------------------|------------|
| Verification Method: | Inspection |
|-----------------------------|------------|

Procedure Number:

Configuration:

Cycle Time:

Verification Discussion/Results:

Up to this time all sources were less than 10 μ curies of FE55 and all individuals directly handling the source container were briefed on the hazards of the source.


 _____ 5/30/97
 ACIS Cognizant Engineer Date