



Notes:

- a) Temperatures shown are for the external mounting surface
- b) Programmed slew rate of interface plate is 0.5C/minute
- c) Dwells start when the ref. temperature is within 2C of target
- d) Except as shown, instrument operates (and is monitored) in transitions

CRaTER Thermal-Vacuum Temperature Profile  
32-06005.0101

Rev. 01 7/18/07  
Rev. 02 7/20/07  
Rev. 03 8/7/07  
Rev. 04 9/21/07