

CRaTER Verification Procedure/Report

Title: Multipaction and Corona

Document 32-06031.06 Rev 01

Reference: 431-SPEC-000008 Rev C, section 3.5
Requirement: High voltage circuits shall be immune to corona in nominal vacuum

Procedure: Demonstrate normal instrument operation in TV testing
(e.g.: through Short Form Functional)

Results:

S/N: _____

Passed/Failed: _____

Comments:

Performed by: _____

R&QA: _____

Date: _____

Date: _____